

Fraunhofer

TESTED® DEVICE

RK Rose+Krieger GmbH Multilift II-400-Clean

Report No. RO 1706-917

Particle Emission

Qualification Certificate

This is to certify that the product mentioned above, provided by

RK Rose+Krieger GmbH Minden, Germany

has been awarded a Fraunhofer certificate TESTED DEVICE bearing the report number RO 1706-917.

The Multilift II-400-Clean was assessed in compliance with ISO 14644-14. When operated under the specified test conditions, it is suitable for use in cleanrooms fullfilling the specifications of the following Air Cleanliness Class according to ISO 14644-1:

Test parameter	Air Cleanliness Class
Travel length: s = 400 mm Velocity: v = 8 mm/s Test load: 40 kg	4
Overall result	4

This document only applies to the named product in its original state and is valid for a period of 5 years from the date the first document was issued. The document can be verified under

www.tested-device.com.

Detailed information and parameters of the test environment can be found in the Fraunhofer IPA test report.

RO 1706-917

Report No. first document

Report No. current document

Stuttgart, February 22, 2018

Place, date of first document issued

Place, current date

r > r

Dr.-Ing. Frank Bürger, Project Manager Fraunhofer IPA

