

TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD  
ZERTIFIKAT ◆ CERTIFICATE ◆ 認證證書 ◆ CERTIFICADO ◆ CERTIFICAT



Product Service

# CERTIFICATE

No. B 012026 0502 Rev. 00

**Holder of Certificate:** **RK ROSE+KRIEGER GmbH**  
**Verbindungs- und Pos.systeme**  
Potsdamer Str. 9  
32423 Minden  
GERMANY

**Production Facility(ies):** 040216  
**Certification Mark:**



**Product:** **General Medical Devices**  
**Lifting column with spindle drive**

**Model(s):** **RKPowerlift M**

**Parameters:** Rated input (TRAFO): 100 / 115 / 120 / 230 / 240 Vac; 75 VA  
Rated input (SMPS): 100 – 240 Vac; 2.5 - 1.0 A  
Rated frequency: 50 / 60 Hz  
Rated input (DC): 24 Vdc (external power supply)  
Operating time: 15% (1.5 min. ON / 8.5 min. OFF)  
Lifting force: 1200 / 1500 / 3000 N

**Brand Name:** RK Rose+Krieger  
**Tested according to:** EN 60601-1:2006/A1:2013

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

**Test report no.:** 713157637  
**Valid until:** 2025-02-09

**Date,** 2020-02-21

  
( Torsten Zimmer )