



Ref. Certif. No.

DE 3 - 41262

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

General Medical Devices
Lifting column with spindle drive

Name and address of the applicant

RK ROSE+KRIEGER GmbH
Verbindungs- und Pos.systeme
Potsdamer Str. 9
32423 Minden
GERMANY

Name and address of the manufacturer

RK ROSE+KRIEGER GmbH Verbindungs- und Pos.systeme
Potsdamer Str. 9, 32423 Minden, GERMANY

Name and address of the factory

Phoenix Mecano Kecskemét Kft.
Szent István krt. 24., 6000 Kecskemét, HUNGARY

Ratings and principal characteristics

Rated input (TRAFO): 100 / 115 / 120 / 230 / 240 Vac; 75 VA
Rated input (SMPS): 100 – 240 Vac; 2.5 - 1.0 A
Rated frequency: 50 / 60 Hz
Rated input (DC): 24 Vdc (external power supply)
Operating time: 15% (1.5 min. ON / 8.5 min. OFF)
Lifting force: 1200 / 1500 / 3000 N

Trade mark (if any)

RK Rose+Krieger

Model/type Ref.

RKPowerlift M

Additional information (if necessary)

Certificate DE 3 - 40330 issued on 2015-06-08 is replaced by this version due to technical changes.

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005
IEC 60601-1:2005/AMD1:2012

as shown in the Test Report Ref. No. which forms part of this certificate

713157637

This CB Test Certificate is issued by the National Certification Body

CB 012026 0503 Rev. 00

Date, 2020-03-18

(Paul Bielawski)



Product Service

Page 1 of 1

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany